Radical-7™

Signal Extraction Pulse CO-Oximeter

Monochrome and Color Display

Service Manual

SET® Rainbow™

The Radical-7 Pulse CO-Oximeter Service Manual is intended to provide the necessary information for proper servicing of all models of the Radical-7 Pulse Oximetry and CO-Oximetry systems. This manual is structured to support troubleshooting to the assembly or module level. This manual does not provide instructions for troubleshooting to the printed circuit board component level. There may be information provided in this manual that is not relevant for your system. Do not service the Radical-7 pulse oximeter without completely reading and understanding these instructions. Service training provided by Masimo is required prior to utilization of this manual.

NOTICE:
Purchase or possession of this device does not carry any express or implied license to use this device with replacement parts which would, alone or in combination with this device, fall within the scope of one of the patents relating to this device.

CAUTION:
Federal law (U.S.) restricts this device to sale by or on the order of a physician.

Masimo Corporation
40 Parker
Irvine, CA 92618
USA
Tel.: 949-297-7000
Fax.: 949-297-7001
www.masimo.com

CE 0123

EU Authorized Representative for Masimo Corporation:

EC REP

MDSS GmbH
Schiffgraben 41
30175 Hannover, Germany
Tel.: +49-511-62 62 86 30
Fax.: +49-511-62 62 86 33

CLASSIFIED c UL us

MEDICAL ELECTRICAL EQUIPMENT WITH RESPECT TO ELECTRIC SHOCK, FIRE AND MECHANICAL HAZARDS ONLY IN ACCORDANCE WITH UL 60601-1/CAN/CSA C22.2 No. 601.1

Covered by one or more of the following U.S. Patents: RE38,492, RE38,476, 7,221,971, 7,215,986, 7,215,984, 7,186,966, 6,979,812, 6,861,639, 6,850,787, 6,826,419, 6,816,741, 6,745,060, 6,699,194, 6,684,090, 6,654,624, 6,650,917, 6,643,530, 6,606,511, 6,515,273, 6,501,975, 6,463,311, 6,430,525, 6,388,240, 6,360,114, 6,263,222, 6,236,872, 6,229,856, 6,157,850, 6,067,462, 6,011,986, 6,002,952, 5,919,134, 5,769,785, 5,758,644, 5,685,299, 5,632,272, 5,490,505, 5,482,036, international equivalents, or one or more of the patents referenced at www.masimo.com/patents.htm Products containing Satshare® feature are also covered by U.S. Patent 6,770,028. Other patents pending.

© 2007 Masimo Corporation. LNOP, Masimo, Radical, the Radical logo, Rainbow, Satshare, SET, Signal IQ, and SpCO are federally registered trademarks of Masimo Corporation. FlashApp, PVI, Radical-7, RDS, SpMet and Signal Extraction Pulse CO-Oximeter are trademarks of Masimo Corporation.

Table of Contents

Page

1 Overview 1-1

1.1 About this Manual 1-1
1.2 Warnings, Cautions and Notes 1-1

2 Maintenance 2-1

2.1 Introduction 2-1
2.2 Cleaning 2-1

3 Battery Operation and Maintenance 3-1

3.1 Overview 3-1
3.2 Battery Charging 3-1
3.3 Replacing the Handheld Battery 3-2
3.4 Replacing the Docking Station Battery 3-2
3.5 Replacing the Fuses 3-3

4 Software 4-1

4.1 Overview 4-1
4.2 Software Download Procedure 4-1

5 Performance Verification 5-1

5.1 Overview 5-1
5.2 Power-On Self-Test 5-1
5.3 Key Press Button Test 5-1
5.4 Alarm Limit Test 5-1
5.5 Display Contrast Test on Monochrome only 5-2
5.6 Testing with Masimo SET Tester 5-2
5.7 %SpCO (Carboxyhemoglobin) Testing (If Applicable) 5-3
5.8 %SpMET (Methemoglobin) Testing (If Applicable) 5-3
5.9 PVI (Pleth Variability Index) Testing (If Applicable) 5-4
5.10 Nurse Call Test 5-4
5.11 Analog Output Test 5-5
5.12 Serial Port 5-5
5.13 Battery Test 5-6

6 Repair 6-1

6.1 Safety Precautions: 6-1
6.2 General Procedures 6-1
6.3 Troubleshooting 6-1
6.4 SatShare Diagnostics 6-7

7 Repair and Returns 7-1

7.1 Repair Policy 7-1
7.2 Return Procedure 7-1

8 Analog Output/Nurse Call Test 8-1

9 Docking Station Power Supply Power Measurement 9-1

10 Portable Functional Block Diagram 10-1
11 Docking Station Functional Block Diagram 11-1
12 Radical-7 Isolation Block Diagram 12-1
13 Portable Wiring Interconnect 13-1
14 Docking Station Wiring Interconnect 14-1
15 Radical-7 Color Parts List 15-1
16 Radical-7 Monochrome Parts List 16-1
17 Docking Station Parts List (RDS-1, RDS-1B and RDS-3) 17-1
18 RDS2 Parts List 18-1

1 Overview

1.1 About this Manual

This manual explains how to service the Radical-7 Pulse CO-Oximeter. Important safety information relating to general use of the Radical-7 Pulse CO-Oximeter appears in the Radical-7 CO-Oximeter Operator's Manual. Other important safety information is located throughout this manual where appropriate.

1.2 Warnings, Cautions and Notes

Please read and follow any warnings, cautions and notes presented throughout this manual. An explanation of these labels is as follows:

A **warning** is provided when actions may result in a serious outcome (i.e., injury, serious adverse affect, death) to the patient or user. Look for text in a gray shaded box.

Sample of Warning:

WARNING: THIS IS A SAMPLE OF A WARNING STATEMENT.

A **caution** is given when any special care is to be exercised by the patient or user to avoid injury to the patient, damage to this device or damage to other property.

Sample of Caution:

CAUTION: This is a sample of a Caution statement.

A **note** is provided when extra general information is applicable.

Sample of Note:

Note: This is a sample of a Note statement.

2 Maintenance

2.1 Introduction

2.1.1 This chapter covers how to test the operation of the Radical-7 and SatShare interface (if applicable), how to properly clean the Radical-7 pulse oximeter, how to recharge and replace the batteries, how to replace the fuses, and how to obtain service under normal operation. No internal adjustment or recalibration is required. Service should be done by qualified personnel only. Safety checks should be performed at regular intervals or in accordance with local and governmental regulations.

WARNING: ELECTRICAL SHOCK AND FLAMMABILITY HAZARD. BEFORE CLEANING THE OXIMETER, ALWAYS TURN THE UNIT OFF AND DISCONNECT THE POWER CORD FROM THE BACK OF THE UNIT.

2.2 Cleaning

2.2.1 To clean the display panel use a cotton swab moistened with 70% isopropyl alcohol and gently wipe the panel.

2.2.2 To clean the outer surface of the oximeter, use a soft cloth moistened with a 70% isopropyl alcohol. Do not allow liquids to enter the interior of the instrument.

2.2.3 To decontaminate unit, refer to Section 7.2.2.

CAUTIONS:

- Do not autoclave, pressure sterilize, or gas sterilize this oximeter.
- Do not soak or immerse the monitor in any liquid.
- Use the cleaning solution sparingly. Allow unit to air dry. Excessive solution can flow into the monitor and cause damage to internal components.
- Do not touch, press, or rub the display panels with abrasive cleaning compounds, instruments, brushes, rough-surface materials, or bring them into contact with anything that could scratch the panel.
- Do not use petroleum-based or acetone solutions, or other harsh solvents, to clean the oximeter. These substances attack the device's materials and device failure can result.

3 Battery Operation and Maintenance

3.1 Overview

3.1.1 The Radical-7 Handheld includes a 2.0 Amp-hr Nickel Metal Hydride (NiMH) battery. The Radical-7 Docking Station may include the optional 6.5 Amp/hr NiMH battery.

3.1.2 Before using the Radical-7 as a handheld or transport monitor, the Handheld battery and the optional Docking Station battery need to be fully charged.

3.2 Battery Charging

3.2.1 To charge the battery(s), attach the Handheld unit to the Docking Station.

3.2.2 Ensure that AC power is attached to the Docking Station.

Note: Some older units are configured with an AC power switch. Turn the AC power switch to the 'I' position.

3.2.3 The battery charging LED indicator(s) on the Docking Station will initially flash, then remain illuminated while the battery(s) is (are) charging. A prolonged flashing battery charging LED indicates that the internal battery temperature exceeds recommended operating conditions for proper battery charging. Proper battery charging will proceed when the temperature returns to recommended operating conditions.

3.2.4 The Handheld battery requires approximately 2 to 3 hours for charging. The optional Docking Station battery requires approximately 6 hours for charging.

3.2.5 When the battery charging LED indicators turn off, additional trickle charging may occur to complete charging. Although battery charging can occur while the Handheld is docked and powered on, most efficient charge times are achieved with the Handheld unit turned off.

3.2.6 It is advisable to completely discharge and fully recharge the batteries periodically to maintain optimum battery health. Refer to the Battery Deep Discharge function as described in the Radical-7 Operators Manual, Section 4, Service. See CAUTIONS: on next page.

CAUTIONS:

- If the Radical-7 has not been charged for 1 month or more, then recharge the battery prior to use.
- At low battery warning, connect the Radical-7 to AC power to prevent interruption of power.
- Additional information on the Radical-7 batteries may be found in the Radical-7 Operators Manual.

3.3 Replacing the Handheld Battery

3.3.1 Turn the Radical-7 Handheld off and remove the patient cable connection. Detach the Radical-7 Handheld from the Docking Station (if docked).

3.3.2 Place the Handheld face down on a soft surface such that no cosmetic damage occurs to the unit.

3.3.3 Loosen the closure screw on the battery, or on color models, the battery compartment door, and remove the battery.

3.3.4 Position the replacement battery into the Handheld, bottom first. On color models, the battery needs to be plugged in / unplugged.

3.3.5 Replace the enclosure (if applicable) and tighten the battery screw until it is secure. If there is a grasp loop, make sure it is pointing towards the top of the Handheld and lays flat into the plastic housing.

3.3.6 Place Handheld into Docking Station. Turn on power and if necessary, charge battery according to Section 9 of the Radical-7 Operator's Manual: Battery Operation and Maintenance.

3.4 Replacing the Docking Station Battery

CAUTION: Do not over-tighten screws. Tighten the screws throughout the Radical-7 Handheld and Docking Station to 80 in.- oz.

Note: Docking Station Battery can only be installed in RDS-1B or RDS-3B type replacement units.

Note: It will be necessary to perform safety testing after replacement.

3.4.1 With the AC power cord disconnected, remove the Handheld from the Docking Station.

3.4.2 Lay the Docking Station on its side on a soft surface, with the screw holes facing up

3.4.3 Remove the five (5) screws from the unit using a #1 Phillips screw driver.

3.4.4 Carefully turn the Docking Station over to allow the screws to fall out of the unit in a manner that allows the technician to catch the screws.

3.4.5 Lay the Docking Station on its feet and remove the top plastic cover.

3.4.6 For units with a battery installed, disconnect the battery connector (J21) from the Docking Station system board. Remove the screw that holds the Docking Station System board in place. Lift this board to allow access to the Battery. Remove the battery.

3.4.7 Carefully install the replacement battery into its compartment.

3.4.8 Reinstall the screw to secure the Docking Station system board. Attach the battery connector.

3.4.9 Carefully install the top plastic cover. Make sure no internal cables are pinched between top and bottom case.

3.4.10 Turn the unit over, and re-install the screws.

3.4.11 Attach the Handheld onto the Docking Station.

3.4.12 If necessary, charge the battery before using the Docking Station according to Section 3.2 of this manual.

CAUTION: Follow local governing guidelines for proper disposal of internal batteries. Do not incinerate.

3.5 Replacing the Fuses

3.5.1 Disconnect unit from AC power.

3.5.2 Use a small flat-blade screwdriver and gently pry loose the fuse holder from the power entry module.

3.5.3 Note how the fuses are placed in the fuse holder for installation of the new fuses.

3.5.4 To remove the fuses from the fuse holder, use the edge of the screwdriver blade to pry against the bottom of the metal portion of the fuse.

3.5.5 Place fuses as needed (1 Amp, Metric, fast acting, 5x20mm, 250V) in the fuse holder, observing the proper orientation.

3.5.6 Slide the fuse holder back into the power entry module and press firmly to make sure it is firmly inserted.

3.5.7 The unit is ready to be reconnected to AC power.

Note: If the fuses blow after replacement, troubleshoot the unit.

WARNING: FIRE HAZARD: TO PROTECT AGAINST FIRE HAZARD, REPLACE ONLY WITH FUSES OF THE SAME TYPE, CURRENT RATING, AND VOLTAGE RATING.

4 Software

4.1 Overview

4.1.1 The Radical-7 product is configured to allow software downloads directly from a PC. The Radical-7 product has three different boards that utilize software that may require upgrade. The Pulse Oximeter Board, Handheld System Board and Docking Station System Board all operate with software that is independent of each other. The software must be compatible between the boards for the Radical-7 product to operate properly. The boards communicate with each other through on board connections - in the case of the Oximeter Board to Handheld System Board. For the Handheld to Docking Station communication, this is accomplished through a Contact Pin/Interface Cable.

4.1.2 When software upgrades are affected, it is imperative that FRM-1312 is completed and returned to Masimo to ensure each unit is tracked with the software version of each unit. This form can be found in this manual, in the contents of the software disk and can also be acquired from Masimo if necessary by contacting Masimo using the information in Section 7.2.4.

Note: The Pulse Oximeter Board (MX-1) is not upgradeable using the PC_FlashApp Program. Only the Docking Station and Portable System Board are upgradeable in this fashion. The upgrade of the MX board requires the use of a special field upgrade tool.

4.2 Software Download Procedure

Note: No software is installed in the RDS-2 and the unit does not require programming.

4.2.1 On the PC, open PC_FlashApp.exe from your hard drive. Masimo software can be obtained by Authorized Service providers by calling Masimo Technical Support Within the United States at (800) 326-4890, option #2, or From Outside the United States at country code (+1) 949-297-7498 option #2.

4.2.2 Set the default “Target”, “Port” and “Baud” parameters as shown in the table below.

Programming Software	Target	Port	Baud[1]
PC_FlashApp	Portable	COM1	57600
PC_FlashApp	Docking Station	COM1	57600

4.2.3 For programming the Radical-7 Handheld System Board, perform the following steps:

4.2.4 Connect a Docking Station to AC outlet and turn the AC Switch to ON (if applicable).

4.2.5 Connect a straight through serial cable between the PC and the Docking Station.

4.2.6 Install the Radical-7 Handheld to the Docking Station.

4.2.7 Turn off the Handheld, if it is on.

4.2.8 Start the FlashApp program. Click Browse and select the proper Image file to be loaded. It will be listed under the monitor's PSA subdirectory.

- If loading on a Radical-7C "Color" the proper file will be named PPSA.ROM.
- If loading on a Radical-7 "Monochrome" the proper file will be named PPSAMX.ROM.

4.2.9 Select the proper Port (usually COM1), and Select Portable (Blue) as the Target.

4.2.10 Press the "Green" Program button. A 'status message' of "waiting for docking station" will be displayed.

4.2.11 Turn on the Radical-7 Handheld.

4.2.12 The green Program button will turn off, and the red Stop button will illuminate. Wait until the green button turns on again indicating a successful transfer.

4.2.13 If the message on the Status window is "Complete – Success", continue with the next step. If programming is not successful, go back to 4.2.7 and repeat. If this cycle is repeated 3 times without success, troubleshooting is required.

4.2.14 Turn off the Handheld.

[1] If a unit does not program with 57,600, set baud rate to 9,600.

4.2.15 To program Radical-7 Docking Station, perform the following steps:

4.2.16 Connect a Docking Station to AC outlet and turn the switch to ON position (if applicable).

4.2.17 Connect a serial cable between computer and the Docking Station.

4.2.18 Install a known good Radical-7 Handheld to the Docking Station.

4.2.19 Turn off the Handheld, if it is on.

4.2.20 Start the FlashApp program. Click Browse and select the proper Image file to be loaded. It will usually be listed under a SDSSA subdirectory.

4.2.21 Select the proper Port (usually COM1), and Select Docking Station as the Target.

4.2.22 Press the “Green” Program button. A ‘status message’ of “waiting for docking station” will be displayed.

4.2.23 Turn on the Handheld.

4.2.24 The green Program button will turn off, and the red Stop button will illuminate. Wait until the green button turns on again indicating a successful transfer. .

4.2.25 If the message on the Status window is “Complete – Success”, continue with next step. If the message is “Complete – Fail”, go back to 4.2.19 and repeat. If this cycle is repeated 3 times without success, troubleshooting is required.

4.2.26 To check software version numbers, perform the following steps:

4.2.27 Turn off the Handheld.

4.2.28 Turn on the Handheld.

4.2.29 Press the Menu key.

4.2.30 Press the Down key and select “About”.

4.2.31 Verify that the Handheld (Handheld, Masimo SET) and Docking Station (D-Station) software version(s) are the intended version. The Handheld must be docked to the Docking Station in order to see the version of the Docking Station.

4.2.32 Record software version in FRM-1312 and notify Masimo as per Section 4.1.2.

5 Performance Verification

Note: Under normal operation, no internal adjustment or recalibration is required.

5.1 Overview

5.1.1 To test the performance of the Radical-7 pulse oximeter following repairs or during routine maintenance, follow the procedure outlined in this section. If the Radical-7 fails any of the described tests, discontinue its use and correct the problem before returning the unit back to the user.

5.1.2 Before performing the following tests place the Radical-7 Handheld into the Docking Station, connect the Radical-7 to AC power and fully charge the Radical-7 Handheld battery (the AC mains switch must be on). Disconnect any patient cables or pulse oximetry sensors, as well as SatShare, serial or Analog Output/Nurse Call cables from the instrument.

5.2 Power-On Self-Test

5.2.1 Connect the unit to AC power and verify that the AC Power Indicator LED is illuminated. If it is not, verify that the AC mains switch is on.

5.2.2 Turn the unit on by depressing the Power/Standby Button. Within 5 seconds all available LEDs are momentarily illuminated, a 1-second beep tone sounds, and the Masimo SET logo is displayed.

5.2.3 The Docking Indicator LED, AC Power Indicator LED, Handheld Battery Charge LED and Docking Station Charge LED (if Docking Station is equipped with optional battery) are illuminated and the Radical-7 begins normal operation.

5.3 Key Press Button Test

5.3.1 With the exception of the Power/Standby Button, press each soft key button and verify that the Radical-7 acknowledges each key-press with an audible beep tone or by indicating an associated change on the display.

5.4 Alarm Limit Test

5.4.1 With the monitor turned on, select the Menu Access key and enter the Alarm menu. Change the desired parameter, below, to a value two points below the currently selected value, and accept the change.

5.4.2 Verify that the newly set parameter is shown on the Saturation Alarm Limit Display, next to the SpO_2 or pulse rate measurement display.

5.4.3 Return the Alarm parameter to its original setting.

5.4.4 Perform steps 5.4.1 through 5.4.3 with the High Saturation Alarm parameter.

5.4.5 Perform steps 5.4.1 through 5.4.3 with the Low Saturation Alarm parameter.

5.4.6 Perform steps 5.4.1 through 5.4.3 with the High Pulse Rate Alarm parameter.

5.4.7 Perform steps 5.4.1 through 5.4.3 with the Low Pulse Rate Alarm parameter.

5.4.8 Reset the alarm limits again to the original settings.

5.5 Display Contrast Test on Monochrome only

5.5.1 With the monitor turned on, select the Menu key and enter the Display menu. Change the Contrast parameter by scrolling through the contrast settings.

5.5.2 Return the Contrast setting to the original value, or a value that allows maximum viewing contrast.

5.5.3 Exit the Menu system and press and hold down the Backlight/Contrast button for several seconds. The display will scroll again through all the contrast settings.

5.5.4 Release the Backlight/Contrast button again when the display shows maximum viewing contrast.

5.6 Testing with Masimo SET Tester

5.6.1 Turn the Radical-7 off and then on again.

5.6.2 Set the alarm limits to:

	% SpO_2	Pulse Rate
High	100	140
Low	90	50

5.6.3 Connect the Masimo tester to the Radical-7. A patient cable LNOP to Red PC cable adapter (4', 8' or 12') may be required to connect a modular tester to the Radical-7.

5.6.4 Verify that within 20 seconds a plethysmographic and a Signal IQ waveform displays.

5.6.5 Verify that the SpO_2 measurement is between 78% and 84%. (with tester values of 81% +/- 3%)

5.6.6 Verify that the pulse rate measurement is between 59 bpm and 63 bpm (with tester values of 61 +/- 1 bpm).

5.6.7 Verify that an audible alarm occurs and that the SpO_2 measurement and the low SpO_2 alarm are flashing

5.6.8 Press the Alarm Silence button once and verify that the alarm is silenced.

5.6.9 Wait 120 seconds and verify that the alarm silence times out and the audible alarm is activated again. Disable the Alarm before continuing.

5.6.10 Press the Increase Loudness button several times and verify that the loudness of the pulse beep tone increases.

5.6.11 Press the Decrease Loudness button and verify that the loudness of the pulse beep tone decreases. Press the Decrease Loudness button and verify that the loudness of the pulse beep tone can be turned off.

5.7 %SpCO (Carboxyhemoglobin) Testing (If Applicable)

Note: This is performed as a functional test only.

5.7.1 Connect a Masimo SET Rainbow sensor to the Radical-7 Handheld.

5.7.2 Insert the non-dominate hand's ring finger into the finger sensor.

5.7.3 Under %SpCO, a healthy, non-smoking individual would have a typical reading between 0 – 5 on the display. (These numbers are for reference only and should not be used for diagnostic purposes)

5.8 %SpMET (Methemoglobin) Testing (If Applicable)

Note: This is performed as a functional test only.

5.8.1 Connect a Masimo SET Rainbow sensor to the Radical-7 Handheld.

5.8.2 Insert the non-dominant hand's ring finger into the finger sensor.

5.8.3 Under %SpMET, a healthy, non-smoking individual would have a typical reading between 0-2% on the display. (These numbers are for reference only and should not be used for diagnostic purposes)

5.9 PVI (Pleth Variability Index) Testing (If Applicable)

Note: This is performed as a functional test only.

5.9.1 Connect a Red Masimo SET Rainbow sensor to the Radical-7 Handheld.

5.9.2 Insert the non-dominant hand's ring finger into the finger sensor.

5.9.3 PVI will fluctuate depending on motion, perfusion, physiology and other factors, so there is no 'normal' value. If a number is displayed, then the PVI feature is functioning properly.

5.10 Nurse Call Test

Note: Refer to Section 8 in this manual for connector pin outs noted in Sections 5.10

5.10.1 Disconnect the patient cable or the Masimo SET Tester from the Radical-7 and turn the instrument on. Ensure that there are no audible alarms and that the audible alarms are not silenced.

5.10.2 Connect the common lead of a digital multimeter to pin 12 (Nurse Call - Common) of the Analog Output/Nurse Call Connector (see Service Manual, Troubleshooting, Section 8, Analog Output/Nurse Call Test) on the Radical-7. Connect the positive lead of the multimeter to pin 6 (Nurse Call - Normally Open) of the Analog Output/Nurse Call Connector and measure that the resistance is greater than 1 MΩ (open circuit).

5.10.3 Trigger an alarm on the monitor (e.g. by disconnecting a sensor after it was measuring data) and verify the resistance is < 20Ω.

5.11 Analog Output Test

5.11.1 Disconnect all patient cables and sensors from the Radical-7. Turn the Radical-7 off and then on again.

5.11.2 Connect the common lead of a multimeter to the pin 2 (Ground) of the Analog Output/Nurse Call Connector. Connect the positive lead of the multimeter to pin 9 (Analog 1) of Analog Output/Nurse Call Connector.

5.11.3 Enter the menu system and set the "Output", "Analog 1" to "0v Output". Verify that the voltmeter measures a voltage of approximately between –10mV and 100mV.

5.11.4 Set “Analog 1” to 1v Output”. Verify that the voltmeter measures a voltage between 0.90V to 1.10V.

5.11.5 Reset “Analog1” to “Spo2 0 – 100%”.

5.11.6 Repeat steps 5.8.3 and 5.8.4 with the positive lead of the multimeter connected to pin 15 (Analog2) and set “Analog2” to “0V Output” and “1V Output”.

5.11.7 Reset “Analog2” to “Pulse Rate”.

5.12 Serial Port

5.12.1 The serial port outputs ASCII data using standard RS-232 protocol via a straight through serial cable. The Handheld must be docked with AC power applied.

5.12.2 Validation of the port can be done by connecting a serial printer or using a PC with Hyper Terminal and the following setup:

5.12.3 Communication Parameters

baud rate	9600 bps
data bits	8
stop bit	1
parity	none
hand shaking	none

5.12.4 With the Radical-7 connected as per the above setup, attach a patient cable and sensor to the Radical-7. From the Output menu, select ASCII 1 and the printer (or PC monitor) will print in the following format every second.

```
05/15/07 08:12:21 SN=0000070986 SPO2=---% BPM=--- PI=--.--%
SPCO=--.-% SPMET=--.-% DESAT=-- PIDELTA=+-- PVI=---
ALARM=0000 EXC=000000
```

5.13 Battery Test

5.13.1 Fully charge the Radical-7 Handheld Battery by placing the Handheld into the Docking Station and connecting the AC power (the AC mains switch must be on).

5.13.2 Verify that the green Handheld Battery Indicator LED is illuminated.

5.13.3 When the Radical-7 is fully charged the green Handheld Battery Indicator turns off.

5.13.4 Turn the Radical-7 on and verify that the Handheld Battery indicator shows a full charge.

5.13.5 Fully charge the Radical-7 Docking Station Battery by placing the Handheld into the Docking Station and connecting the AC power (the AC mains switch must be on).

5.13.6 Verify that the green Handheld Battery and blue Connection Indicator LEDs, on the Docking Station, are illuminated.

5.13.7 When the Radical-7 is fully charged the green Docking Station Battery Indicator turns off.

5.13.8 Turn the Radical-7 on and verify that the Docking Station Battery indicator shows a full charge.

6 Repair

6.1 Safety Precautions:

WARNING: DO NOT DISASSEMBLE OR ASSEMBLE THE UNIT WITH THE AC POWER CORD ATTACHED.

WARNING: ENSURE THE UNIT HAS BEEN CLEANED PER THE CLEANING INSTRUCTIONS.

6.2 General Procedures

6.2.1 Examine the unit prior to disassembly or troubleshooting and inspect for cosmetic damage. External damage may be helpful in determining the root cause of the failure.

6.2.2 Note any markings or labeling the end user may have placed on the unit that may need to be duplicated or transferred in the event the housing or ancillary components are to be replaced.

6.2.3 Make note of the unit serial number.

6.2.4 Make note of the software versions in the event the Oximeter board or system boards are replaced, as these boards require software updates to make them compatible with the end users version of software.

6.2.5 Contact Masimo Corporation using the information in Section 7.2.4 to research the warranty history of the unit to determine if the unit will be covered under warranty.

WARNING: A HI-POT SAFETY TEST IS REQUIRED IN THE EVENT THE DOCKING STATION HAS BEEN DISASSEMBLED.

6.3 Troubleshooting

6.3.1 The troubleshooting procedures in this manual instruct the technician to isolate the failure down to the sub-assembly level. The intention is to have the Unit Under Test (UUT) repaired with minimal steps. A safety analyzer or Hi-pot/Current Leakage test unit will be needed in the event the Docking Station is disassembled. The Masimo SET Tester will be needed to perform the Performance Verification in Section 5 upon completion of repairs.

Tools required are as follows:

Phillips head screwdriver
Multimeter
10mm wrench or small pliers

6.3.2 The troubleshooting steps below are presented in order of the most probable cause first. Proceed to the next step only if unit continues to fail.

6.3.3 Always note the software versions of all three boards prior to repair.

6.3.4 All units must have Performance Verification in Section 5 performed after repairs have been affected.

6.3.5 All Docking Stations that have been opened must be safety tested after reassembly.

6.3.6 Service Messages

6.3.6.1 R2.0.x.x – When a fault is detected, the unit either watchdogs or displays “Service Required XXX” (where “XXX” indicates one of the symptoms below), and generates high priority alarm sound. In either case the fault is latched until power is cycled or the alarm silence key is pressed.

6.3.6.2 R2.1.x.x – When a fault is detected, the unit either watchdogs or displays “XXX” and generates high priority alarm sound. When the fault triggers watchdog alarm, it is latched until the power is cycled or the alarm silence key is pressed. But when the unit displays “XXX”, it also resets itself to clear the fault. If the fault is one-time event, software clears the fault and continues to operate normally. If the fault is permanent, the Radical-7 will display “XXX”, reset, display “XXX”, reset…

Symptom	Cause and Corrective Action Steps
• **001**	• **SRAM Memory Failure** 1. Replace Portable System Board
• **002**	• **Flash Memory Failure** 1. Replace Portable System Board
• **003**	• **Low Speed Bus Timeout** 1. Replace Portable System Board
• **004**	• **Display or Low Speed Bus Failure** 1. Replace Portable System Board

Symptom	Cause and Corrective Action Steps
	2. Replace Display
• **005**	• **Dual UART Failed at Power Up** 1. Replace Portable System Board
• **006**	• **Oximeter Board Failure** 1. Replace Oximeter Board
• **007**	• **Oximeter MX Board Communication Error** 1. Verify Oximeter Board (SET) Software is correct a. If version is incorrect or not displayed, return unit to an authorized service center or use a field upgrade tool to reload the oximeter board software. b. If version is correct, replace Oximeter Board. c. Replace Portable System Board
• **008**	• **Real Time Clock Failure** 1. Replace Portable System Board
• **009**	• **Watchdog Failure** 1. Replace Portable System Board
• **010**	• **Docking Station Watchdog Failure** 1. Replace Docking Station System Board. 2. Replace Portable System Board
• **200**	• **Invalid SatShare coefficients** 1. Replace Docking Station System Board.

Symptom	Cause and Corrective Action Steps
• **All Flashing LED's** 1. Flashing alternates between all LED's and AC Power LED. 2. Refer to Radical-7 Operators Manual for LED status Parameters.	• **Communication error between Portable and Docking Station** 1. Verify software versions of Portable and Docking Station System Boards using the Portable menu a. If versions are incorrect or not displayed, reload System Application Software for that device utilizing procedures outlined in the Software

Symptom	Cause and Corrective Action Steps
	Download Procedure. 2. Replace The Interface Cable 3. Replace the Portable System Board 4. Replace the Docking Station System Board
• **No LED'S Illuminated** 1. Both Portable and Docking Station are turned on. 2. Refer to Radical-7 Operators Manual for LED status Parameters.	• **Power fault** 1. Replace external Power Cable 2. Check fuses at AC Input Filter 3. Check 12VDC power (Sections 7 & 8) 4. Replace Docking Station System Board
• **No Battery Icon Displayed**	• **NiMH Battery Fault** 1. Replace NiMH battery 2. Replace Portable System Board
• **Battery LED Not Illuminated**	1. No fault, battery is fully charged 2. Check 12VDC power (Sections 7 & 8) 3. Replace Docking Station System Board
• **Battery LED Flashing**	1. Ensure Service Mode – Deep Discharge is not enabled 2. Allow NiMH Battery to cool down 3. Verify Fan operation 4. Replace NiMH Battery 5. Replace Docking Station System Board.
• **Alarm LED does not Illuminate** 1. Unit must be in alarm state 2. Audible alarm can be heard	1. Replace Docking Station System Board 2. Replace LED Cable.
• **Alarm Condition But No Audible Indication.** 1. Unit must be in alarm condition. 2. Should hear sound upon button depression.	• **Alarm Fault** 1. Replace Speaker 2. Replace Portable System Board
• **Display Missing Segments**	1. Replace LCD 2. Replace Portable System Board
• **Display Backlight Will Not Illuminate**	1. Replace Portable System Board 2. Replace Display
• **Button Function Error**	1. Replace Keypad

Symptom	Cause and Corrective Action Steps
1. Any front panel button affected 2. Buttons do not activate feature 3. Buttons hard to press	2. Replace Portable System Board
• **Rightmost Touch Key Difficult to Depress**	1. Open unit. If Swivel Connector flex cable is interfering with button, replace Swivel Connector 2. Replace Keypad.
• **"No Sensor", "Defective Sensor" or "Sensor Off" Displayed**	1. Replace sensor. 2. Replace patient cable. 3. Replace oximeter board.
• **Display Does Not Rotate Automatically**	1. Ensure Docking Station is powered up 2. Ensure Power Save feature is not enabled 3. Replace Docking Station System Board
• **Display Does Not Rotate Manually**	1. Flex interference? Replace Swivel Connector. 2. Replace Keypad 3. Replace Portable System Board
• **Docking Station Fails When Portable is Docked**	1. Ensure software of Handheld and Docking Station is compatible 2. Replace Portable System Board 3. Replace Docking Station System Board 4. Replace Interface Cable
• **Fan Does Not Function** 1. Ensure AC power is on	1. Replace Fan 2. Replace Docking Station System Board
• **Portable Will Not Power Up When Undocked** 1. If Clicking is heard when On/Off button is depressed, perform Step 3.	1. Check NiMH Battery charge 2. Replace Battery if charging is not successful 3. Replace Portable System Board.
• **Portable will not Power Up When Docked** 1. Docking Station is functioning 2. Portable powers up undocked	1. Replace Interface Cable 2. Replace Docking Station System Board
• **Portable Will Not Turn Off**	1. Replace Portable System Board 2. Adjust or replace Keypad

Symptom	Cause and Corrective Action Steps
• **Portable Will Not Undock** • **From Docking Station**	1. Ensure Portable Lock is not in 2. Latch failure, gently pry latch down; once removed, replace latch and/or striker
• **Watchdog Alarm** 1. Other than "Service" failures	1. Replace Portable System Board

WARNING: IF SYMPTOMS ARE PRESENT AND ARE NOT CORRECTABLE, DO NOT USE MALFUNTIONING COMPONENT.

6.4 SatShare Diagnostics

CAUTION: As SatShare Diagnostics simulates physiological information (SpO_2 and Pulse Rate values), it should be performed with caution, especially in areas with central monitoring. The attending nurse and supervisor should always be made aware that SatShare Diagnostic testing is being performed.

6.4.1 If the SpO_2 and pulse rate values displayed on a multi-parameter monitor is in question, perform the following steps.

6.4.1.1 Press Menu – General – SatShare Numbers – Yes.

6.4.2 With the Handheld docked and AC power applied, disconnect the patient cable from Radical-7. Press the following key sequence: Menu – About – More – 2 – 1 – 1 (see Service menu for number key assignments). **Five (5) different SatShare values can be selected by pressing the Select key.** The selected value will be displayed on the multi-parameter monitor if the unit is working properly. The SatShare Diagnostics feature times out if a key is not pressed within 1 minute.

Note: During transient situations, the SpO_2 and pulse rate values on the Radical-7 and multi-parameter monitor may differ by more than the specified tolerance. The averaging time set on the multi-parameter monitor may also affect the delay during transient situations.

7 Repair and Returns

7.1 Repair Policy

7.1.1 Masimo or an authorized Service Department must perform warranty repair and service. Do not use malfunctioning equipment. The product must be repaired.

7.1.2 Please clean contaminated/dirty equipment before returning. Make sure it is fully dry before packing the equipment.

7.1.3 To return the Radical-7 unit for service, please follow the Return Procedure outlined in this section.

7.2 Return Procedure

7.2.1 Please clean contaminated/dirty equipment before returning and make sure it is completely dry before packing the equipment.

7.2.2 Recommended decontamination procedure for Radical-7 equipment that has the potential for being exposed to blood or other potentially infectious materials:

7.2.2.1 Put on Lab coat or other protective garment.

7.2.2.2 Put on gloves.

7.2.2.3 Place equipment on the designated decontamination table.

7.2.2.4 Apply cleaning agent to paper towel. Clean with 70% Isopropyl alcohol wipes, or 1:10 bleach to water solution. Wipe down all aspects of the equipment.

7.2.2.5 Let the equipment air dry completely.

7.2.2.6 Remove gloves.

7.2.2.7 Move equipment to the area designated for equipment that has been through the decontamination process.

7.2.2.8 Dispose of all contaminated products properly.

7.2.3 Package the equipment securely – in the original shipping container if possible – and enclose the following information and items:

7.2.3.1 Call Masimo From within the United States at (800) 326-4890, option #2, or from outside the United States at country code (+1) 949-297-7498, option #2 and ask for an RMA number.

7.2.3.2 Place a copy of the supplied RMA inside the box, and write the RMA number on the outside of the shipping container

7.2.3.3 A letter or email describing in detail any symptoms or difficulties experienced with the pulse oximeter.

7.2.3.4 Warranty information – a copy of the invoice or other applicable documentation must be included

7.2.3.5 Purchase order number to cover repair if the oximeter is not under warranty, or for tracking purposes if the warranty is in effect

7.2.3.6 Ship-to and bill-to information

7.2.3.7 Person (name, telephone/Telex/fax number, email and country) to contact for any questions about the repairs

7.2.3.8 A document stating the oximeter has been decontaminated for blood borne pathogens

7.2.4 Return the pulse oximeter and a copy of the RMA to the following shipping address:

Masimo Corporation
40 Parker
Irvine, California
USA 92618

(Domestic Toll Free): 800-326-4890, option #2
(International): 949-297-7498, option #2
FAX: 949-297-7499
Email: **tech@masimo.com**

8 Analog Output/Nurse Call Test



9 Docking Station Power Supply Power Measurement

Part Number 20241 and 20364



Part Number 20293



Docking Station System Board Power Measurement



10 Portable Functional Block Diagram



11 Docking Station Functional Block Diagram



12 Radical-7 Isolation Block Diagram




Displa
y
J6
Portable System
Board
J7
J3
MX- x
Board
J1
Front Panel
Patient
Connector
Speaker
J1
J16
J9
NiMH
Battery
Pack
J17
J8
J19
Connection
to
Docking Station

14 Docking Station Wiring Interconnect



See Page 11, Operators Manual, for illustration of the back panel

15 Radical-7 Color Parts List



1. Shell, Front, Radical-7 Color
2. Lens, Radical-7 Color
3. Keypad, Radical-7 Color
4. LCD Module, 480 x 272, Color TFT
5. Bracket, LCD, Radical-7 Color
6. Radical-7c User Interface circuit board
7. Screw #4-40 x 3/8” Phillips, pan head machine, 18-8 SS
8. Radical-7 Color system circuit board
9. Standoff, Hex 3/16”, #4-40 IT, 5/16” L ACI, RoHS Compliant
10. Screw, #4-40 x 3/16” Phillips, pan head machine, 18-8 SS
11. MX-1, Standard
12. Speaker, Radical, Reinforcement
13. Spring, Speaker, Radical-7 Color
14. Assembly, FIC, Radical-7 Color
15. Shell, Back, Radical-7 Color
16. Screw, #4-40 x 3/4” Phillips, pan head machine, 18-8 SS
17. Latch, Radical-7 Color
18. Battery Pack, NiMH, AA, 4.8V, 2000mAH
19. Door, Battery, Radical-7 Color
20. Label, Radical-7 Color, HH, SN, 4-UP, MX-1
21. Battery, Lithium, 12.5MM coin, 3v, 48mAH
22. Tape, Polyimide, Acrylic Adhesive, 0.75” w, .0025” THK
23. Label, Radical-7 Color Screen, connector with notch

WARNING: Disassembly or repair by authorized Service Personnel only.

16 Radical-7 Monochrome Parts List



1. Latch
2. Top Shell
3. Swivel Connector
4. Leash Pins
5. Speaker Gasket
6. Speaker
7. Keypad Adhesive
8. Keypad
9. Display.
10. Display Retainer
11. MX-x Standoff Mounting Hardware (4 pcs)
12. Handheld System Board
13. Portable System Board Mounting Hardware (2 pcs)85
14. Standoff (4 pcs)
15. MX-xx mounting Hardware (4 pcs)
16. MX-1, MX-3
17. Back Shell
18. Plastics Hardware (6 pcs)
19. NiMH Battery Pack
20. Radical-7 Logo Label
21. Lithium Coin Cell
22. Foam Pad

Portable Board – MX-x board compatibility Table (lower case x indicates compatibility):

Portable System Board	MX-1[2]
20755	20699
Radical 7 Portable System Board Re-order Part Number	20841
MX-1 Board Re-order Part Number	2143

WARNING: Disassembly or repair by authorized Service Personnel only.

[2] The Oximeter Board re-order part number is different than the part number label on the board. The re-order part number indicates that the board has been tested and programmed. The portable Systems Board and the MX-x Circuit Board must be compatible. Re-order the part number as indicated in the table above.

17 Docking Station Parts List (RDS-1, RDS-1B and RDS-3)



Note: Only RDS-1B configured Docking Stations will contain the 20284 Optional Battery Pack.

1. Front Shell
2. LED Status Label
3. Radical Logo Label
4. LED Status Cable
5. Interface Cable
6. Interface Cable Hardware (4 pcs [2x2])
7. Bottom Shell
8. Equi-potential Hardware (4 different pcs)
9. Plastics Hardware (5 pcs)
10. AC Filter Module Fuses (2 pcs)
11. AC Power Entry Module
12. AC Filter Module Bracket
13. Fan
14. AC Power Entry Module Bracket Hardware (5 pcs [1x2x2])
15. AC Cable Insulator
16. Power Supply
17. **20179** AC Power Cable or **20320** (required with 20293 Power Supply)
18. Insulator
19. DC Power Cable
20. Top Shell
21. Docking Station System Board (2 types)
    - **RDS-1 and RDS-1B (All ports available – P1, P2 and P3)**
    - **RDS-3 (Only ports P1 and P2 available)**
22. Standoff
23. Docking Station System Board Mounting Screw
24. No Implied License Label
25. Certification Label – (Note: Port P3 Closed
    - **RDS-1 and RDS-1B (All ports open)**
    - **RDS-3 (P3 closed)**
26. Power Supply Hardware (4 pcs)
27. Front Shell Mounting Hardware (4 pcs [2x2])
28. Docking Station Battery Pack (Optional)

WARNING: Disassembly or repair by authorized Service Personnel only.

18 RDS2 Parts List



1. Front Shell
2. LED Status Label
3. Radical Logo Label
4. LED Status Cable
5. Interface Cable
6. Interface Cable Hardware (4 pcs [2x2])
7. Bottom Shell
8. Equipotential Hardware (4 different pcs)
9. Plastics Hardware (5 pcs)
10. AC Filter Module Fuses (2 pcs)
11. AC Power Entry Module
12. AC Filter Module Bracket
13. AC Power Entry Module Bracket Hardware (5 pcs [1x2x2])
14. AC Cable Insulator
15. Power Supply
16. **20179** AC Power Cable or **20320** (required with 20293 Power Supply)
17. Insulator
18. DC Power Cable
19. Top Shell
20. RDS2 System Board
21. Standoff
22. Docking Station System Board Mounting Screw
23. No Implied License Label
24. Power Supply Hardware (4 pcs)
25. Certification Label – **(Port P1, P2 and P3 closed)**
26. Front Shell Mounting Hardware (4 pcs [2x2])

WARNING: Disassembly or repair by authorized Service Personnel only.